



SGS

EU Type Examination Certificate Number: **0120/SGS0306**

F&F Filipowski s.j.

Ul. Konstytucyjna 79/81
95-200 Pabianice
Poland

Instrument Identification:
LE-03 Series

Instrument Traceable Number
0120/SGS0306

Polyphase, Active Import/Export (kWh), Electricity Meter

has been assessed and certified as meeting the requirements of

EU Directive 2014/32/EU

on Measuring Instruments Annex II, Module B

It is certified that the manufacturer's technical design and specimen for the above instrument has been examined and, based on the evidence submitted, it is considered that the instrument conforms to the requirements of Annex V of EU Directive 2014/32/EU

This certificate must be used in conjunction with a certificate covering the product verification as required in Annex II, Module D or Annex II, Module F

This certificate is valid until 14th March 2027
Issue 1

Certification is based on report number(s) SHES151200818401 Dated 17th March 2017
EMA234429/1
EMA238979

Authorised Signature

SGS United Kingdom Limited, Notified Body 0120
Unit 202B Worle Parkway, Weston-super-Mare, BS22 6WA, UK
t +44 (0)1934 522917 f +44 (0)1934 522137 www.sgs.com

Contact Address
SGS United Kingdom Ltd, Units 12A & 12B, South Industrial Estate, Bowburn, Durham, DH6 5AD, UK
t +44 (0)191 377 2000 f +44 (0)191 377 2020 www.sgs.com



This document is issued, on the Client's behalf, by the Company under its General Conditions of Service printed overleaf. The Client's attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein.


MID-B-06E Rev 9

Any other holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Clients instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents.

EU Type Examination Cert.

SGSPAPER
18103276



	EU-Type Examination Certificate Number:	
	0120/SGS0306	
	Issue Number: 1	Dated: 26 th May 2017

1. Technical Data

Manufacturer	F&F Filipowski s.j.
Meter Type(s)	LE-03R LE-03M LE-03MW
Voltage Rating (U_n)	3x230/400V
Current Rating ($I_{min} - I_{ref} (I_{max})$)	0.25-5(30)A, 0.25-5(32)A, 0.25-5(40)A, 0.25-5(45)A, 0.25-5(50)A, 0.25-5(80)A,
Frequency (F_n)	50Hz
Active Accuracy Class (kWh)	A or B (kWh)
Type of circuit	3P4W
Temperature Range	-25°C to +55°C
Software Version No's.	LE-03R – V1.01 LE-03M – V1.01 LE-03MW – V1.01
Checksum	91E4
Identification Location	LCD & Nameplate
Bill Of Materials No.'s	LE-03R: D519032 LE-03M: D519033 LE-03MW: D519024
IP Rating	IP51
Insulation Protective Class	Class II
LED Pulse Constant	1000imp/kWh
Impulse Voltage Rating	6kV
AC Voltage Rating	4kV
Main Cover Sealing Type	Wire & Crimp
Integrity of meter	Inaccessible without breaking seals
Intended Location of the Meter	Indoor
Type of Register	LCD
Terminal Arrangement(s)	DIN
Location of Manufacturers Name & Address	User Manual

SGS

EU-Type Examination Certificate Number:

0120/SGS0306


Issue Number: 1

Dated: 26th May 2017


2. Photographs of Meter and Sealing Plan



Terminal cover seals








Main cover seal

	EU-Type Examination Certificate Number:	
	0120/SGS0306	
	Issue Number: 1	Dated: 26 th May 2017




3. Examples of Nameplates



«F&F»[®] **LE-03R** EN50470-1/3

CE **M17** **0120**



 kWh
 -25°C to 55°C
 3X230/400V
 0,25-5(80)A
 S.W.1.01
 50Hz




1000imp/kWh 1000imp/kvarh  



«F&F»[®] **LE-03M** EN50470-1/3

CE **M17** **0120**



 kWh
 -25°C to 55°C
 3X230/400V
 0,25-5(80)A
 S.W.1.01
 50Hz

1000imp/kWh 1000imp/kvarh  


«F&F»[®] **LE-03MW** EN50470-1/3

CE **M17** **0120**



 kWh
 -25°C to 55°C
 3X230/400V
 0,25-5(80)A
 S.W.1.01
 50Hz

1000imp/kWh 1000imp/kvarh  

4. Influence factors for temperature, frequency and voltage

		Influence Factors for Temperature. Frequency & Voltage					
Current	PF Cos	-25°C	-10°C	5°C	30°C	40°C	55°C
I _{min}	1.0	0.38	0.25	0.13	0.08	0.17	0.29
I _{tr}	1.0	0.38	0.27	0.14	0.07	0.16	0.31
10I _{tr}	1.0	0.37	0.25	0.13	0.08	0.18	0.33
I _{max}	1.0	0.32	0.22	0.16	0.06	0.15	0.28
I _{tr}	0.5ind	0.38	0.27	0.13	0.06	0.14	0.31
10I _{tr}	0.5ind	0.36	0.24	0.11	0.09	0.19	0.35
I _{max}	0.5ind	0.30	0.21	0.14	0.05	0.15	0.31
I _{tr}	0.8cap	0.37	0.26	0.14	0.07	0.17	0.33
10I _{tr}	0.8cap	0.35	0.23	0.11	0.11	0.20	0.36
I _{max}	0.8cap	0.30	0.21	0.13	0.07	0.16	0.31
L1					0.00	0.00	0.00
I _{tr}	1.0	0.39	0.26	0.12	0.07	0.17	0.26
10I _{tr}	1.0	0.39	0.25	0.11	0.09	0.18	0.29
I _{max}	1.0	0.30	0.21	0.13	0.06	0.15	0.26
I _{tr}	0.5ind	0.39	0.26	0.12	0.09	0.15	0.30
10I _{tr}	0.5ind	0.38	0.25	0.11	0.10	0.19	0.33
I _{max}	0.5ind	0.30	0.21	0.13	0.06	0.15	0.28
L2							
I _{tr}	1.0	0.52	0.37	0.20	0.10	0.24	0.46
10I _{tr}	1.0	0.51	0.35	0.17	0.12	0.26	0.49
I _{max}	1.0	0.45	0.33	0.21	0.08	0.21	0.42
I _{tr}	0.5ind	0.54	0.36	0.20	0.08	0.22	0.44
10I _{tr}	0.5ind	0.51	0.35	0.17	0.11	0.25	0.47
I _{max}	0.5ind	0.47	0.34	0.22	0.08	0.20	0.41
L3							
I _{tr}	1.0	0.17	0.12	0.08	0.07	0.12	0.28
10I _{tr}	1.0	0.15	0.12	0.06	0.05	0.13	0.27
I _{max}	1.0	0.09	0.09	0.07	0.04	0.10	0.23
I _{tr}	0.5ind	0.16	0.12	0.06	0.05	0.09	0.26
10I _{tr}	0.5ind	0.12	0.11	0.06	0.08	0.13	0.28
I _{max}	0.5ind	0.06	0.08	0.02	0.05	0.11	0.24


	EU-Type Examination Certificate Number:	
	0120/SGS0306	
	Issue Number: 1	Dated: 26 th May 2017

During the type approval examination the influence factors for temperature, frequency and voltage are determined per load point. The table above represents the sum of the square values per load, determined via the following formula:-

$$\delta e (T, U, f) = \sqrt{(\delta e^2 (T, I, \cos\varphi) + \delta e^2 (U, I, \cos\varphi) + \delta e^2 (f, I, \cos\varphi))}$$

where

- $\delta e(T, I, \cos\varphi)$ = Additional error due to variation of the temperature at the same load
- $\delta e(U, I, \cos\varphi)$ = Additional error due to variation of the voltage at the same load
- $\delta e(f, I, \cos\varphi)$ = Additional error due to variation of the frequency at the same load


	EU-Type Examination Certificate Number:	
	0120/SGS0306	
	Issue Number: 1	Dated: 26 th May 2017

5. Annex of Variants

Product Variant Identification Details:

Type Designation	Description of meter
LE-03R	3x230/400V, 0.25-5(x), energy measurement(kWh & kVarh), IR
LE-03M	3x230/400V, 0.25-5(x), energy measurement(kWh & kVarh), IR, RS485(Modbus Protocol)
LE-03MW	3x230/400V, 0.25-5(x), energy measurement(kWh & kVarh), IR, RS485(Modbus Protocol), Multi Tariff

Modifications to the meter(s) described according to approval No.**0120/SGS0306** must be notified to the issuing body to confirm the meter(s) continuing compliance to the relevant pattern approval standard(s).

	EU-Type Examination Certificate Number:	
	0120/SGS0306	
	Issue Number: 1	Dated: 26 th May 2017

6. Document Revision History

Issue	Date	Comments
1	26/05/2017	Initial Issue

“This document is issued by the Company subject to its General Conditions of Service available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at www.sgs.com/terms_e-document.htm. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company’s findings at the time of its intervention only and within the limits of Client’s instructions, if any. The Company’s sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.”
 “Unless otherwise stated the results shown in this test certificate refer only to the sample(s) tested *and such sample(s) are retained for 28 days only.*”

**** End of Document ****
